

COVID 19 ANTIGEN CARD

Rapid Visual Test for the Qualitative Detection of Covid-19 (SARS-COV-2) antigen in Human nasopharyngeal and nasal swab specimens

INTRODUCTION

Novel corona virus infection SARS CoV-2 [COVID-19] has spread to more than 203 countries of various regions including Africa, America, Europe, South East Asia and Western Pacific. The WHO had declared COVID-19 as the global public health emergency and subsequently as pandemic because of its worldwide spread. It is now one of the top-priority pathogens to be dealt with, because of high transmissibility, severe illness and associated mortality, wide geographical spread, lack of control measures with knowledge gaps in veterinary and human epidemiology, immunity and pathogenesis. The quick detection of cases and isolating them has become critical to contain it. Whereas molecular diagnostic tests were rapidly developed, antigen based tests are of utmost requirement for early diagnosis of infection. Antigen is generally detectable in upper respiratory specimen during acute phase of infection.

INTENDED USE

COVID 19 Antigen Card is a visual, rapid and sensitive immunoassay for the qualitative detection of COVID-19 (SARS-CoV-2) Antigen in human nasopharyngeal, nasal swab specimen. **The test is a screening test only and is an aid for the early diagnosis of SARS-CoV-2 infection in patient with clinical symptoms.** More specific and alternate diagnostic method should also be performed for confirmation of SARS-CoV-2 infection.

PRINCIPLE

Covid 19 Antigen Card is an immunoassay based on "Sandwich" principle. The test contains two pre-coated lines, "C" Control line and "T" test line. SARS-CoV-2 specific Antibody is immobilized on nitrocellulose membrane at "T" Line. Colloidal gold is conjugated to another antibody specific for SARS-CoV-2 Antigen. The test sample treated with extraction buffer when added to the device, interacts with the Colloidal gold Conjugate. If the specimen contains a novel Coronaviruses antigen, it will form a complex with the Colloidal gold-labelled with SARS-CoV-2 antibody. The complex migrates through the nitrocellulose strip by capillary action. When the complex meets the line of corresponding immobilized antibody, the complex is trapped forming a pink purple band which confirms a reactive test result. Absence of a colored band in the test region indicates a non-reactive test result. A red procedural control line should always develop at "C" region to indicate that the test has been performed properly.

KIT CONTENTS

COVID 19 Antigen Test Cards (with desiccant)
Sample Extraction Buffer
Sample Extraction Tube
Sample Dropper
Sterile Nasal Swabs
Sample Extraction Tube Stand
Instruction Manual

25 TESTS

25 Nos.
1 Vial
25 Nos.
25 Nos.
25 Nos.
1 No.
1 No.

PACKSIZE

25 Tests, 50 Tests, 100 Tests & 200 Tests

KIT STORAGE AND STABILITY

The kit should be stored at 2-30°C in the cool and driest area available. Expiry date on the kit indicates the date beyond which kit and its components should not be used. Covid 19 Antigen Card should not be frozen and must be protected from exposure to humidity.

MATERIAL REQUIRED BUT NOT PROVIDED

The kit contains all the items required to perform this test. However, a timer or stop watch is required.

WARNING FOR USERS

CAUTION: ALL THE SAMPLES TO BE TESTED SHOULD BE HANDLED AS THOUGH CAPABLE OF TRANSMITTING INFECTION.

- The use of Disposable Gloves, glasses and proper protecting clothing is STRONGLY RECOMMENDED while running the test.
- This test detects the presence of covid 19 antigen in the human nasopharyngeal / nasal swab specimens only. **Do not use Human Whole blood (EDTA Anticoagulated), plasma sera, urine and saliva as specimen for this test.**
- Deviation in the test procedure may adversely affect test performance and/or produce invalid results and hence adherence to test procedure and precautions given in the product insert shall be strictly followed.
- In case there is a cut or wound in hand, DO NOT PERFORM THE TEST.
- Do not smoke, drink or eat in areas where specimens or kit reagents are being handled.
- Do not pipette by mouth.
- Tests are for in vitro diagnostic use only and should be run by competent trained person only.
- Dispose off all specimens and materials used to perform the test as bio-hazard waste. Laboratory chemical and biohazard wastes must be handled and discarded in accordance with all local, state, and national regulations for Covid 19.
- Spills should be decontaminated promptly with Sodium Hypochlorite or any other suitable disinfectant.
- Wash hands thoroughly with soap or any suitable detergent, after the use of the kit. In case of needle prick or other skin puncture or wounds, wash the hands with excess of water and soap.
- Do not perform the test in area with strong air flow i.e. under fan or strong air conditioning.
- Observe established precautions against microbiological hazards throughout sample collection and testing procedures.
- All specimens to be tested should be handled in accordance with the laboratory practice for infectious diseases.

PRECAUTION

Optimal assay performance requires strict adherence to the assay procedure described in the manual.

- Bring all the reagents to room temperature (20-30°C) before use.
- Do not use kit components beyond the expiration date, which is printed on the kit.
- Use a separate swab and Sample extraction tube for each specimen and then discard it as biohazardous waste.
- Do not re-use the test kit.
- Do not use the test kit if the pouch is damaged or the seal is broken.
- Do not mix or use the sample extraction buffer of different lots.
- Do not smoke, drink or eat while handling specimen.
- Avoid microbial and cross contamination of reagents.
- Ensure that the sample extraction buffer is added in the tube upto the mark provided (0.4).**

BEFORE YOU START

The Sample Extraction Buffer provided in the kit has closed nozzle and screw cap with pin (outside). Before using Sample Extraction Buffer, keep the vial vertically straight and tap down gently on the working platform, so that Sample Extraction Buffer comes down at the bottom of the vial. **To orifice/puncture the closed nozzle, follow the instruction as illustrated on back side of this instruction manual (Point No. 5).**

SPECIMEN COLLECTION, SAMPLE PREPARATION AND STORAGE

Human nasopharyngeal swab specimens should be used for the test.

- Take out required no. of Sample Extraction Tube from the pouch and place in the Sample Extraction Tube Stand provided with the kit.
- Label them with patient I.D.
- Add Sample Extraction Buffer from the vial in Sample Extraction Tube upto the mark provided (0.4 ml).
- Nasopharyngeal Swab specimen collection:** Take the nasal swab provided in the kit and carefully insert into the nostril presenting the most secretion on visual inspection. Keep the swab near the septum floor of the nose by gently pushing the swab into the posterior nasopharynx. Gently rotate the swab several times against the nasal wall and remove it from the nasopharynx.

Care should be taken to collect mucus free swab specimen.

5. Sample preparation:

- Insert the collected specimen swab (as mentioned above) into the Sample Extraction tube containing Sample Extraction Buffer.

- Swirl the swab 10 times, and squeeze the swab head against the tube wall of the extraction tube to release the antigen in the buffer to remove as much liquid as possible from the swab. **Dispose off swab according to biohazard waste disposal method.**
- Tightly close the cap of the tube and mix the contents by gentle shaking (5 times).
- Sample Storage:** It is recommended to test the patient specimens immediately after collection. Specimens collected in Sample extraction buffer may be stored at 2-8°C for upto 4 hours or at room temperature upto 1 hour.

TEST PROCEDURE

- Take out the COVID 19 Antigen test card from the foil pouch prior to use and place it on a flat and dry surface.
- Transfer 5 drops (150 µl) of the extracted sample (from step 2 "Sample Preparation") using sample dropper provided into the sample well of the test card. Dispose off the sample dropper immediately considering it to be potentially biohazardous.
- Allow the reaction to occur for 20 minutes.
Read the result at 20 minutes. Do not read the results after 30 minutes. Results beyond 30 minutes may give false results.

NOTE: Discard the device immediately after reading the results, considering it to be infectious.

INTERPRETATION OF TEST RESULTS

REACTIVE

Appearance of two distinct pink line in the control region 'C' and test region 'T' indicates that the specimen is reactive for Covid 19 antigen.

NON-REACTIVE

Appearance of one distinct pink line in the control region 'C' only and no line in the test region 'T' indicates that the specimen is non-reactive for Covid 19 antigen.

INVALID

If no "C" Control line appears after the test is completed, with/ without "T" test line; the test indicates ERROR. **The left over collected specimen (from sample collection and preparation step) shall be Centrifuged at 10,000 rpm for 15 minutes and re-run the test using fresh new device.** The invalid result may be obtained due to following reasons:

- Improper storage at temperature other than the recommended temperature.
- Wrong sample collection and / or test procedure.
- Long atmospheric exposure of the test device after opening the pouch.

PERFORMANCE CHARACTERISTICS

The performance of Covid 19 Antigen Card has been evaluated by ICMR approved Lab for COVID-19 using RT-PCR confirmed positive and negative samples. The result obtained of 3 different lots are as follows:

Lot Details	No. of Positive samples	No. of Negative samples	Sensitivity	Specificity
LOT 1	20	10	95%	100%
LOT 2	20	10	95%	100%
LOT 3	20	10	95%	100%

LIMITATIONS OF THE TEST

- The kit works best when used with fresh specimens. Specimens containing mucus/ particulates can result in improper flow of extracted sample & hence making the interpretation of results difficult.
- Potentially interfering substances like Nasal Sprays, nasal drops, Homeopathic allergic release medicines, anti-inflammatory medicines, anti-viral drugs, Biotin etc. may lead to False positive results.
- Any deviation from test procedure may lead to erratic results.
- This is only a screening test and should not be used as the sole criteria for the diagnosis of Covid-19 infection.**
- All samples detected non-reactive must be confirmed by using RT-PCR. Therefore for a definitive diagnosis & more accuracy of immune status, the patient's clinical history and symptomatology should be considered in order to obtain a confirmation of Covid-19 infection. A negative result at any time does not preclude the possibility of an early infection of Covid-19 infection. The results should be reported only after complying with above procedure.**
- This is a Qualitative test and therefore neither determine Covid 19 Antigen quantitative value nor the rate of SARS-CoV-2 antigen concentration.
- For more accuracy of immune status, additional follow-up testing using other laboratory methods is recommended.
- The test result must always be evaluated with other data available to the physician.
- A negative result may occur if the concentration of antigen in a specimen is below the detection limit of the test or if the specimen was collected or transported improperly. **Therefore a negative test result does not eliminate the possibility of SARS-CoV-2 infection, and should be confirmed by RT-PCR.**
- Positive test results do not rule out co-infections with other pathogens.

LIMITED EXPRESSED WARRANTY DISCLAIMER

The manufacturer limits the warranty to the test kit, as much as that the test kit will function as an *in vitro* diagnostic assay within the limitations and specifications as described in the product instruction manual, when used strictly in accordance with the instructions contained therein. The manufacturer disclaims any warranty expressed or implied including such expressed or implied warranty with respect to merchantability, fitness for use or implied utility for any purpose. The manufacturer's liability is limited to either replacement of the product or refund of the purchase price of the product and in no case liable to claim of any kind for an amount greater than the purchase price of the goods in respect of which damages are likely to be claimed. The manufacturer shall not be liable to the purchaser or third parties for any injury, damage or economic loss, howsoever caused by the product in the use or in the application there of.

BIBLIOGRAPHY

- Schoy A., Anantharajah A., Bodeus M., Kabamba-Mukadi B., Verroken A., Rodriguez-Villalobos H.: Low performance of rapid antigen detection test as frontline testing for COVID-19 diagnosis. J. Clin Virol. 2020 Aug; 129: 104455. doi:10.1016/j.jcv.2020.104455
- Nicholas C.G., Pons-Salort M., White P.J., N., Ferguson N. M.: Comparison of molecular testing strategies a mathematical modeling study. THE LANCET Infectious Diseases. DOI: http://doi.org/10.1016
- Patterson E.I., Prince T., Anderson E.R., Casas-Sanchez A., Smith S.L., Cansado-Utrilla C., Turtle L., Hughes G.L.: Method of inactivation of SARS-CoV-2 for downstream biological assays.
- Che Xi-Y., Hao W., Wang Y., Di B., Xu Yin-C., Feng Ch-S., Wan Zh-Y., Cheng V.C.C., Yuen Kw-Y.: Nucleocapsid Protein as Early Diagnostic Marker for SARS; Emerg Infect Dis. 2004 Nov; 10(11):1947-1949. doi:10.3201/eid1011.040516

REFER THE BACK PAGE OF THE INSTRUCTION MANUAL FOR PICTORIAL REPRESENTATION & BETTER CLARITY OF THE TEST PROCEDURE AND RESULT INTERPRETATION.

in vitro diagnostic reagent, not for medicinal use

Manufactured & Marketed By:

DIAGNOSTIC ENTERPRISES


Plot No.: 26, Indl. Estate, Sector-1, Parwanoo - 173 220, (H.P.)
Phone: 01792-232253 E-mail: de@diagnosticenterprises.com

COVID 19 ANTIGEN CARD

Rapid Visual Test for the Qualitative Detection of Covid-19 (SARS-COV-2) antigen in Human nasopharyngeal and nasal swab specimens

MATERIAL PROVIDED


Individually foil pouched test device with a desiccant.



Sample Extraction Buffer Vial



Sample Extraction Tube




Sample Dropper



Nasopharyngeal Swab



Sample Extraction Tube Stand



Instruction Manual



PREPARATION FOR THE TEST


1) Carefully read the instruction manual for using the Covid 19 Antigen Card Test




2) Bring the required number of test devices, Sample Extraction Tubes and Sample Extraction Buffer vial to room temperature prior to testing.

RT
20-30°C

3) Check the expiry date at the back of test device foil pouch and Sample Extraction Buffer vial label. **Use fresh kit if expiry date has passed.**




4) Take out required no. of Sample Extraction Tube from the pouch, label it with patient ID and place in the Sample Extraction Tube Stand.




5) Make Sample Extraction Buffer vial ready to use



6) Cut open the device foil pouch and check the color of the desiccant pouch, it should be blue in color. **Do not use the device if desiccant is pink in color.**




7) Label the test device with patient's name or identification number.




TEST PROCEDURE

1) SPECIMEN COLLECTION & PREPARATION:


i) Add Sample Extraction Buffer in Sample Extraction Tube upto the mark provided (0.4 ml)




ii) Take nasopharyngeal swab sample




iii) Place patient swab specimen in Sample Extraction tube



iv) Swirl the swab atleast 10 times and squeeze the swab head against the sample extraction tube wall to release the antigen in the sample extraction buffer and discard the swab.



v) Close the cap tightly and mix the contents of the vial by gentle shaking (5 times).




2) RUNNING THE TEST ON THE DEVICE

Take the specimen reagent mixture from the sample extraction tube using sample dropper provided with the kit.




Add 5 drops (150 µl) of the extracted sample using sample dropper into the sample well of the test card.



3) READING TIME

Read result at 20 Minutes




20
minutes

Do not read results after 30 minutes


RESULT INTERPRETATION

REACTIVE for Covid 19 Antigen




Appearance of red coloured line in the control region 'C' and Test region 'T' indicates that the specimen is reactive for Covid 19 Antigen.

INVALID



If neither control line nor test line appears or only test line appears, the test should be treated as Invalid. Repeat the test again.

NON-REACTIVE



Appearance of one red coloured line in the control region 'C' indicates that the specimen is non-reactive for Covid 19 Antigen.